



HONEYCOMB

Honeycomb Biotechnologies
940 Winter St
Waltham, MA 02451
support@honeycomb.bio

14 December, 2021

Certificate of Analysis

Component	Part number	Kit lot #	Date of manufacture	Expiration date	Acceptance Criteria	Result
HIVE scRNAseq Molecular Controls	10369	2134801EDH	12/14/21	06/14/22	Library yield 1-50 ng/ μ L; Product peak 500- 1200 bp	PASS; PASS

Description of Assembly: In-process controls for use with the HIVE scRNAseq Processing Kit.

Storage Temperature: -20°C

This document serves to certify that the product(s) identified above conform to all applicable specifications as evidenced by reports or other documentation on file.

Signature:

X 

Name: Ryan Stipetic

Title: Senior Operations Manager